

Corona Canine & Feline

Coronavirus (CCV, FCoV) Ag Detection kit

For in vitro veterinarian diagnostic use only

Cat.no – 80COV205/80COV250

Instructions for Use

Intended use

ImmunoRun Corona test kit is intended for the detection of specific antigens of Coronavirus (CCV, FCoV) in feces from dogs and cats.

Specifications


Samples	Fresh dog or cat feces		
Time to result	10 minutes		
Storage temp.	2-30°C		
Shelf life	24 months		
*	Sensitivity	Specificity	Accuracy
CCV Ag	99.99%	97.50%	97.67%

*According to internal comparison study 2016/2017.
ELISA (Enzyme-Linked Immunosorbent Assay).

Components of the test kit

Components	5 Tests/ kit 80COV205	50 Tests/kit 80COV250
Test Cassette	5	50
Cotton Swab	5	50
Sample tube	5	50

Components of the test kit



Note: In the reaction field, before starting the test, a green/blue line appears in the control line region. This is used for quality control and will be washed away by the sample liquid during the test.

Precautions



- Do not freeze the kit.
- Do not open or remove the test cassette from its individually sealed pouch until it is to be used.
- Do not use the test if the cassette pouch or the device is damaged.
- Do not touch the exposed membrane in the device window
- Handle and dispose all contaminated materials in accordance with approved sanitary standards for biohazardous waste.
- Components in this kit have been quality control approved as a standard batch unit. Do not mix components from different lot numbers.
- Each component of the kit is intended for a single use only.

STORAGE

- Store at 2-30°C. Avoid exposure to direct sunlight.
- The kit is stable for up to 24 months, do not use beyond the expiration date.

Sample Preparation


- Collect fresh fecal samples using the swab. Insert the tip of the swab into the sample and rotate to collect a thin film of feces. Repeat this sampling in at least **3 different locations** of the fecal sample. Alternatively, insert the swab through the rectal sphincter and gently rotate to recover visible fecal material.
- Make sure the cotton swab is evenly covered with feces.
- Do not collect an excessive amount of feces as this may lead to a false result or test malfunction.**



- Do not collect samples from cat litter or other coarse particles.
- For pooled fecal samples, stir well prior to testing.
- For postponed testing, samples can be stored up to **maximum 6 days at 2 to 8°C**.

Test Instructions

- If stored in refrigeration, allow all kit components and specimen to reach room temperature prior to testing.
- Immediately insert the swab into the sample tube and stir vigorously (approximately 10 seconds) to assure good sample dissolvment. Close the tube and shake briefly.
 - Remove the test cassette from its pouch and place it on a horizontal surface.
 - Use the test cassette **within 60 minutes** of opening the pouch.
- Break off the upper tip of the sample tube by pushing it firmly.
- Add 3 drops of the sample material to the sample well.



Note: If the fluid does not begin to migrate through the test strip, use the cotton swab's upper tip to remove any solid material which may be interfering with the flow.


If still no migration occurs, apply another drop of liquid to the corresponding well.

To detect intermittent shedding and to determine whether an animal is infected with coronavirus, at least **three fecal samples** should be examined at intervals of **one week to one month**.

Test Results

The test results should be read after **10 minutes**.

Positive test result:




Two red bands will appear.

The upper control band (C) confirms that the test is working properly. The bottom test band (T) indicates a positive test result.

**A weak test line should also be considered a positive antigen test result.*

**A weak control line may occur when the applied sample is too concentrated, or as a result of certain components in the sample which interfere with the test run.*


Negative test result:



A lack of a test band (T), while control band (C) is present.

***No antigen detected.**

Invalid test result:



If **no control band (C)** appears the test is considered invalid.

If the **test line is colored brown** by feces, this test is invalid and should be repeated.

Limitations & Troubleshooting

A low incidence of false results can occur.

All results must be considered together with additional clinical information available.

Symbols

	Observe product information		For one-time use only		Storage temperature		Protect against light
	For professional use only		Number of Test/kit		Shelf life		Protect from humidity
			Lot number				Manufacturer

Biogal Galed Labs.

YOUR
DIAGNOSTIC
COMPANION

www.biogal.com

For further information and assistance please contact your local distributor or Biogal Galed Labs. Acs. Ltd. directly by e-mail: info@biogal.com or by tel: 972-4-9898605

Note: Coronaviruses are usually shed periodically. However, there are also animals that permanently shed the virus and thus continuously contaminate the environment.